

细胞培养系列最终质量检查表 Tissue Culture Products Quality Final Check List

目录号 Catalog Number	TCD010060	规格 Specifications	6cm
批号 Lot Number	230702-206BA	有效期 Expiration Date	2026-07-02
生产日期 Manufacture Date	2023-07-02	检验日期 Inspection Date	2023-07-05

检验项目及结果

Inspection Item and Result

检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无油污、破裂、缩水变形现象, 无披锋; 包装完整 No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
性能 Function	密封性 Sealing Property	N/A				
	容量: ±5% Capacity: within ±5%	N/A				
	悬浮型 Suspension type	N/A				
	表面处理 Surface Treatment	C=0	3	0: 1	符合/Meet criteria	合格/OK
	细胞培养: 贴壁率>80% Tissue Culture: Adherence rate >80%	C=0	3	0: 1	符合/Meet criteria	合格/OK
生物性能 Biological Performance	无菌 Sterilized(SAL=10 ⁻⁶)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	热原≤0.03EU/mL Non-Pyrogenic (≤0.03EU/mL)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	细胞毒性 Cytotoxicity	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
最终判断 Final Judgement		接收 Accept			拒收 Reject	
签名 Signature						
检验 Inspected by			审核 Approved by			
张 1 1 1			马 1 1 1			