
Certificate of Conformity

Product: Nunc Centrifuge Tube, 11 ml**Date:** 2020-09-08**Page:** 1**Item number:** 347856**Lot number:** 166290**Date of manufacture:** August 2020**Date of expiry:** August 2025

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation", with SAL 10E-6.

The product is CE marked according to EU directive, 2007/47/EC amending 93/42/EC and 98/79/EC for IVD, Medical device.

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality management system - Requirements" and ISO 13485:2016 "medical devices - Quality management systems - System requirements for regulatory purposes".

Roskilde Site



Emil O'Hanlon Cohrt, PhD

Quality Systems, Support and Monitoring, Team Leader