

细胞培养系列最终质量检查表 Tissue Culture Products Quality Final Check List

目录号 Catalog Number		TCF011050		规格 Specifications		50mL	
批号 Lot Number		230430-600BA		有效期 Expiration Date		2026-04-30	
生产日期 Manufacture Date		2023-04-30		检验日期 Inspection Date		2023-05-03	
检验项目及结果 Inspection Item and Result							
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result		
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac: Re	结果 Result	判断 Judgement	
外观 Appearance	产品无油污、破裂、缩水变形现象, 无披锋; 包装完整 No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria		合格/OK
性能 Function	密封性 Sealing Property	C=0	10	0: 1	符合/Meet criteria	合格/OK	
	容量: ±5% Capacity: within ±5%	C=0	10	0: 1	符合/Meet criteria	合格/OK	
	悬浮型 Suspension type	N/A					
	表面处理 Surface Treatment	C=0	3	0: 1	符合/Meet criteria	合格/OK	
	细胞培养: 贴壁率>80% Tissue Culture: Adherence rate >80%	C=0	3	0: 1	符合/Meet criteria	合格/OK	
生物性能 Biological Performance	无菌 Sterilized(SAL=10 ⁻⁶)	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK	
	热原≤0.03EU/mL Non-Pyrogenic (≤0.03EU/mL)	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK	
	细胞毒性 Cytotoxicity	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK	
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK	
最终判断 Final Judgement		接收 Accept			拒收 Reject		
签名 Signature							
检验 Inspected by				审核 Approved by			
张 1 生				马 菊 菊			