



细胞培养系列最终质量检查表 Tissue Culture Products Quality Final Check List						
目录号 Catalog Number	TCF012050		规格 Specifications	50ml		
批号 Lot Number	230103-312BA		有效期 Expiration Date	2026-01-03		
生产日期 Manufacture Date	2023-01-03		检验日期 Inspection Date	2023-01-06		
检验项目及结果 Inspection Item and Result						
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无油污、破裂、缩水变形现象, 无披锋; 包装完整 No oil, crack, shrinkage deformation, burrs. With complete packaging.		依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5		符合/Meet criteria 合格/OK	
性能 Function	密封性 Sealing Property		N/A			
	容量: ±5% Capacity: within ±5%		C=0	10	0: 1	符合/Meet criteria 合格/OK
	悬浮型 Suspension type		N/A			
	表面处理 Surface Treatment		C=0	3	0: 1	符合/Meet criteria 合格/OK
	细胞培养: 贴壁率>80% Tissue Culture: Adherence rate >80%		C=0	3	0: 1	符合/Meet criteria 合格/OK
生物性能 Biological Performance	无菌 Sterilized(SAL=10 <sup>-6</sup> )		C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: --- 合格/OK
	热原≤0.5EU/mL Non-Pyrogenic (≤0.5EU/mL)		C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: --- 合格/OK
	细胞毒性 Cytotoxicity		C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: --- 合格/OK
	无酶 DNase & RNase Free		C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: --- 合格/OK
最终判断 Final Judgement	接收 Accept				拒收 Reject	
签名 Signature						
检验 Inspected by			审核 Approved by			
张子强			马萧萧			