

离心管 (瓶) 最终质量检查表 Centrifuge Tubes Quality Final Check List

目录号 Catalog Number	CFT011500	规格 Specifications	50mL					
批号 Lot Number	210715-060	有效期 Expiration Date	2024-07-15					
生产日期 Manufacture Date	2021-07-15	检验日期 Inspection Date	2021-07-18					
检验项目及结果 Inspection Item and Result								
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result			
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement		
外观 Appearance	产品无油污、破裂、缩水变形现象, 无披锋; 包装完整 No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK		
性能 Function	离心力测试: 锥底: 12,000×g Centrifuge test: Conical tubes: 12,000×g	C=0	10	0: 1	符合/Meet criteria	合格/OK		
	容量: ±2% Capacity: within ±2%	C=0	10	0: 1	符合/Meet criteria	合格/OK		
	耐高温: 121°C, 20 分钟 High temperature performance: 121°C, 20min	C=0	10	0: 1	符合/Meet criteria	合格/OK		
	冻存测试: -80°C, 24 小时 Freezable: -80°C, 24hour	C=0	10	0: 1	符合/Meet criteria	合格/OK		
	油墨牢固度 Ink fastness	C=0	10	0: 1	符合/Meet criteria	合格/OK		
	密封性测试: Sealability test	C=0	10	0: 1	符合/Meet criteria	合格/OK		
生物性能 Biological Performance	无菌 Sterilized(SAL=10 ⁻⁶)	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK		
	热原≤0.5EU/mL Non-Pyrogenic (≤0.5EU/mL)	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK		
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK		
最终判断 Final Judgement	接收 Accept			拒收 Reject				
签名 Signature								
检验 Inspected by			审核 Approved by					
张 江 伟			马 菊 菊					