

Quality Certificate

Product Name : 150ml Vacuum Filtration

Catalog Number : FPE204150

Batch Number : 211210-053-A

Expiry Date : 20241210

Customer Focus:

At Jet, our customers' satisfaction is exactly what we are pursuing. We have good management and product development teams who have years of industry experience. Our cell and tissue culture professionals and our entrepreneurial spirit make us responsive to customer needs. We welcome special requests and work to provide the best and most personal service in the industry.

Satisfaction Guarantee:

Jet offers a total satisfaction guarantee so you can be confident in your purchasing decision. If for any reason you are not satisfied with the product performance or service provided by Jet, we will either replace or issue a refund for the purchase price of your product.

JET products are manufactured in accordance with ISO 13485 : 2016 & ISO 9001 : 2015 quality management systems in a class 100,000 clean room.

Membrane: PES 0.22 μm

Flow Rate- Flow rate of this product has been tested no less than 150 ml/min with deionized water under -70KPa.

Sterilization - Product labeled as sterile is irradiated and dosimetrically released upon ISO11137 recommended practices in effect at the time of validation. Sterilization records are reviewed and signed off by qualified personnel for product release. JET BIOFIL products labeled sterile meet a minimum requirement of 10^{-6} SAL (Sterility Assurance Level).

Non-Pyrogenic - Products labeled non-pyrogenic have been validated per FDA guidelines on LAL (Limulus Amebocyte Lysate) testing for medical devices and company guidelines. The acceptance level for product is less than 0.5 EU/ml.

DNase & RNase Free - This product has been tested and is free of any detectable DNase / RNase contamination.

Max. Vacuum Pressure - The max. vacuum pressure of this lot product has been validated. The acceptance level for product is not less than 0.09MPa and no any distortion or damaged.

Quality Control Testing - Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.



Quality Manager