

细胞刮刀最终质量检查表 Cell Scraper Quality Final Check List

目录号 Catalog Number	CSC011025	规格 Specifications	25cm			
批号 Lot Number	221107-233-A	有效期 Expiration Date	2025-11-07			
生产日期 Manufacture Date	2022-11-07	检验日期 Inspection Date	2022-11-10			
检验项目及结果 Inspection Item and Result						
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收:退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无油污、破裂、缩水变形现象， 无披锋；包装完整 No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准：GB/T 2828.1-2012 Standard：GB/T2828.1-2012 抽样水平：Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
性能 Function	适配性 Compatibility	C=0	10	0:1	符合/Meet criteria	合格/OK
生物性能 Biological Performance	无菌 Sterilized(SAL=10 ⁻⁶)	C=0	3	0:1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	热原≤0.5EU/mL Non-Pyrogenic (≤0.5EU/mL)	C=0	3	0:1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	无酶 DNase & RNase Free	C=0	3	0:1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -	合格/OK
最终判断 Final Judgement		接收 Accept	√		拒收 Reject	
签名 Signature						
检验 Inspected by				审核 Approved by		
张 江 强				马 菊 菊		