

离心管（瓶）最终质量检查表 Centrifuge Tubes Quality Final Check List

目录号 Catalog Number	CFT921500	规格 Specifications	50mL			
批号 Lot Number	230703-060-D	有效期 Expiration Date	2026-07-03			
生产日期 Manufacture Date	2023-07-03	检验日期 Inspection Date	2023-07-06			
检验项目及结果 Inspection Item and Result						
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收：退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无油污、破裂、缩水变形现象，无 披锋；无缺胶；包装完整 No oil, crack, shrinkage deformation,burrs; No shortage of glue; With complete packaging.	依据标准：GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平：Level II, AQL 1.5(严重缺陷：C=0) Sample Level: Level II, AQL 1.5 (Serious flaws: C=0)			符合/Meet criteria	合格/OK
性能 Function	离心力测试：锥底:20,000×g Centrifuge test : Conical tubes: 20,000×g	C=0	10	0: 1	符合/Meet criteria	合格/OK
	容量：±2% Capacity: within ±2%	C=0	10	0: 1	符合/Meet criteria	合格/OK
	耐高温：121℃，20 分钟 High temperature performance: 121℃, 20min	C=0	10	0: 1	符合/Meet criteria	合格/OK
	冻存测试：-80℃，24 小时 Freezable:-80℃，24hour	C=0	10	0: 1	符合/Meet criteria	合格/OK
	油墨牢固度 Ink fastness	C=0	10	0: 1	符合/Meet criteria	合格/OK
	书写效果 writing effect	C=0	10	0: 1	符合/Meet criteria	合格/OK
	密封性测试（-90kpa,60s）： Sealability test（-90kpa,60s）	C=0	10	0: 1	符合/Meet criteria	合格/OK
生物性能 Biological Performance	无菌 Sterilized(SAL=10 ⁻⁶)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -	合格/OK
	热原≤0.03EU/mL Non-Pyrogenic (≤0.03EU/mL)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -	合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -	合格/OK
最终判断 Final Judgement		接收 Accept √			拒收 Reject	
签名 Signature						
检验 Inspected by				审核 Approved by		
张 江 强				马 菊 菊		