

### 微量离心管最终质量检查表 Micro Centrifuge Tubes Quality Final Check List

<b>目录号</b> Catalog Number	CFT000015	<b>规格</b> Specifications	1.5ml
<b>批号</b> Lot Number	220628-119-CD	<b>有效期</b> Expiration Date	2027-06-28
<b>生产日期</b> Manufacture Date	2022-06-28	<b>检验日期</b> Inspection Date	2022-07-01

**检验项目及结果**

**Inspection Item and Result**

检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收:退 Ac:Re	结果 Result	判断 Judgement
<b>外观</b> Appearance	产品无油污、破裂、缩水变形现象, 无披锋; 包装完整 No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
	离心力测试: 25,000×g Centrifuge test : 25,000×g	C=0	10	0: 1	符合/Meet criteria	合格/OK
<b>性能</b> Function	容量: ±2% Capacity: within ±2%	N/A				
	耐高温: 121°C, 20分钟 High temperature performance: 121°C, 20min	C=0	10	0: 1	符合/Meet criteria	合格/OK
	冻存测试: -80°C, 24小时 Freezable: -80°C, 24hour	C=0	10	0: 1	符合/Meet criteria	合格/OK
	翻盖耐折性 Cover folding resistance	C=0	10	0: 1	符合/Meet criteria	合格/OK
	管体配合及密封性 Tube fit and sealing	C=0	10	0: 1	符合/Meet criteria	合格/OK
<b>生物性能</b> Biological Performance	无菌 Sterilized(SAL=10 <sup>-6</sup> )	N/A				
	热原≤0.5EU/mL Non-Pyrogenic (≤0.5EU/mL)	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK

<b>最终判断</b> Final Judgement	接收 Accept	√	拒收 Reject
--------------------------------	--------------	---	--------------

<b>签名 Signature</b>	
<b>检验 Inspected by</b>	<b>审核 Approved by</b>
张 江 华	马 萧 萧

