

微量离心管最终质量检查表 Micro Centrifuge Tubes Quality Final Check List						
目录号 Catalog Number		CFT002015		规格 Specifications		1.5ml
批号 Lot Number		221115-184-D		有效期 Expiration Date		2027-11-15
生产日期 Manufacture Date		2022-11-15		检验日期 Inspection Date		2022-11-18
检验项目及结果 Inspection Item and Result						
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac: Re	结果 Result	判断 Judgement
外观 Appearance	产品无油污、破裂、缩水变形现象, 无披锋; 包装完整 No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
性能 Function	离心力测试: 25,000×g Centrifuge test : 25,000×g	C=0	10	0: 1	符合/Meet	合格/OK
	容量: ±2% Capacity: within ±2%	N/A				
	耐高温: 121°C, 20 分钟 High temperature performance: 121°C, 20min	C=0	10	0: 1	符合/Meet	合格/OK
	冻存测试: -80°C, 24 小时 Freezable: -80°C, 24hour	C=0	10	0: 1	符合/Meet	合格/OK
	翻盖耐折性 Cover folding resistance	C=0	10	0: 1	符合/Meet	合格/OK
	管体配合及密封性 Tube fit and sealing	C=0	10	0: 1	符合/Meet	合格/OK
生物性能 Biological Performance	无菌 Sterilized(SAL=10 ⁻⁶)	N/A				
	热原≤0.5EU/mL Non-Pyrogenic (≤0.5EU/mL)	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK
最终判断 Final Judgement		接收 Accept			拒收 Reject	
签名 Signature						
检验 Inspected by				审核 Approved by		
张 江 强				马 菊 菊		