

## 微量离心管最终质量检查表 Micro Centrifuge Tubes Quality Final Check List

| 目录号<br>Catalog Number                   | CFT001050                                                                                           | 规格<br>Specifications                                                                                              | 5ml                |                |                                                             |                 |
|-----------------------------------------|-----------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------|--------------------|----------------|-------------------------------------------------------------|-----------------|
| 批号<br>Lot Number                        | 230916-3274-A                                                                                       | 有效期<br>Expiration Date                                                                                            | 2026-09-16         |                |                                                             |                 |
| 生产日期<br>Manufacture Date                | 2023-09-16                                                                                          | 检验日期<br>Inspection Date                                                                                           | 2023-09-19         |                |                                                             |                 |
| 检验项目及结果<br>Inspection Item and Result   |                                                                                                     |                                                                                                                   |                    |                |                                                             |                 |
| 检验项目及标准<br>Inspection Item and Standard |                                                                                                     | 接收水平<br>Acceptance Level                                                                                          |                    |                | 检验结果<br>Inspection Result                                   |                 |
|                                         |                                                                                                     | 检验水平<br>Quality Level                                                                                             | 样品数<br>Sample Size | 收: 退<br>Ac:Re  | 结果<br>Result                                                | 判断<br>Judgement |
| 外观<br>Appearance                        | 产品无油污、破裂、缩水变形现象, 无披锋; 包装完整<br>No oil, crack, shrinkage deformation, burrs. With complete packaging. | 依据标准: GB/T 2828.1-2012<br>Standard: GB/T2828.1-2012<br>抽样水平: Level II, AQL 1.5<br>Sample Level: Level II, AQL 1.5 |                    |                | 符合/Meet criteria                                            | 合格/OK           |
| 性能<br>Function                          | 离心力测试: 25,000×g<br>Centrifuge test : 25,000×g                                                       | C=0                                                                                                               | 10                 | 0: 1           | 符合/Meet criteria                                            | 合格/OK           |
|                                         | 容量: ±2% Capacity: within ±2%                                                                        | C=0                                                                                                               | 10                 | 0: 1           | 符合/Meet criteria                                            | 合格/OK           |
|                                         | 耐高温: 121°C, 20 分钟<br>High temperature performance: 121 °C, 20min                                    | C=0                                                                                                               | 10                 | 0: 1           | 符合/Meet criteria                                            | 合格/OK           |
|                                         | 冻存测试: -80°C, 24 小时<br>Freezable: -80°C, 24hour                                                      | C=0                                                                                                               | 10                 | 0: 1           | 符合/Meet criteria                                            | 合格/OK           |
|                                         | 翻盖耐折性 Cover folding resistance                                                                      | C=0                                                                                                               | 10                 | 0: 1           | 符合/Meet criteria                                            | 合格/OK           |
|                                         | 管体配合及密封性<br>Tube fit and sealing                                                                    | C=0                                                                                                               | 10                 | 0: 1           | 符合/Meet criteria                                            | 合格/OK           |
| 生物性能<br>Biological Performance          | 无菌<br>Sterilized(SAL=10 <sup>-6</sup> )                                                             | C=0                                                                                                               | 3                  | 0: 1           | 样品 Samples: ---<br>阳性对照 Positive: +++<br>阴性对照 Negative: --- | 合格/OK           |
|                                         | 热原≤0.03EU/mL<br>Non-Pyrogenic (≤0.03EU/mL)                                                          | C=0                                                                                                               | 3                  | 0: 1           | 样品 Samples: ---<br>阳性对照 Positive: +++<br>阴性对照 Negative: --- | 合格/OK           |
|                                         | 无酶<br>DNase & RNase Free                                                                            | C=0                                                                                                               | 3                  | 0: 1           | 样品 Samples: ---<br>阳性对照 Positive: +++<br>阴性对照 Negative: --- | 合格/OK           |
| 最终判断 Final Judgement                    |                                                                                                     | 接收 Accept                                                                                                         | √                  |                | 拒收 Reject                                                   |                 |
| 签名 Signature                            |                                                                                                     |                                                                                                                   |                    |                |                                                             |                 |
| 检验 Inspected by                         |                                                                                                     |                                                                                                                   |                    | 审核 Approved by |                                                             |                 |
| 张 江 伟                                   |                                                                                                     |                                                                                                                   |                    | 马 菊 菊          |                                                             |                 |

