

# Quality Certificate

**Product Name :** 125ml Erlenmeyer Flasks

**Catalog Number :** TAB101125

**Batch Number :** 210712-685

**Expiry Date :** 20240712

## Customer Focus:

At Jet, our customers' satisfaction is exactly what we are pursuing. We have good management and product development teams who have years of industry experience. Our cell and tissue culture professionals and our entrepreneurial spirit make us responsive to customer needs. We welcome special requests and work to provide the best and most personal service in the industry.

## Satisfaction Guarantee:

Jet offers a total satisfaction guarantee so you can be confident in your purchasing decision. If for any reason you are not satisfied with the product performance or service provided by Jet, we will either replace or issue a refund for the purchase price of your product.

**JET products are manufactured in accordance with ISO 13485: 2016 & ISO 9001: 2015 quality management systems in a class 100,000 clean room.**

**Materials** - Flask : PETG (Polyethylene Terephthalate Glycol)  
Cap: HDPE (High-density Polyethylene)

**Sterilization** - Product labeled as sterile is irradiated and dosimetrically released upon ISO 11137 recommended practices in effect at the time of validation. Sterilization records are reviewed and signed off by qualified personnel for product release. JET BIOFIL products labeled sterile meet a minimum requirement of 10<sup>-6</sup> SAL (Sterility Assurance Level).

**Non-Pyrogenic** - Products labeled non-pyrogenic have been validated per FDA guidelines on LAL (Limulus Amebocyte Lysate) testing for medical devices and company guidelines. The acceptance level for product is less than 0.5 EU/mL.

**Low temperature test** - This product has been tested is freezable at -80°C for 48hour without distortion.

**DNase & RNase Free** - This product has been tested and is free of any detectable DNase / RNase contamination.

**Cytotoxicity** - Testing is conducted to qualify all material resins using USP and ISO 10993 standards for cytotoxicity and have been shown to be non-toxic.

**Quality Control Testing** - Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

  
**Quality Manager**