

细胞培养系列最终质量检查表 Tissue Culture Products Quality Final Check List

目录号 Catalog Number	TCP001096	规格 Specifications	96 孔
批号 Lot Number	221109-078-F	有效期 Expiration Date	2025-11-09
生产日期 Manufacture Date	2022-11-09	检验日期 Inspection Date	2022-11-12

检验项目及结果

Inspection Item and Result

检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无油污、破裂、缩水变形现象, 无披锋; 包装完整 No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
性能 Function	密封性 Sealing Property	N/A				
	容量: $\pm 5\%$ Capacity: within $\pm 5\%$	N/A				
	悬浮型 Suspension type	C=0	3	0: 1	符合/Meet criteria	合格/OK
	表面处理 Surface Treatment	N/A				
	细胞培养: 贴壁率>80% Tissue Culture: Adherence rate >80%	N/A				
生物性能 Biological Performance	无菌 Sterilized(SAL= 10^{-6})	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK
	热原 $\leq 0.5\text{EU/mL}$ Non-Pyrogenic ($\leq 0.5\text{EU/mL}$)	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK
	细胞毒性 Cytotoxicity	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK
最终判断 Final Judgement		接收 Accept ☒			拒收 Reject	

签名 Signature

检验 Inspected by	审核 Approved by
张 1 强	马 萧 萧

