

移液管最终质量检查表 Serological pipettes Quality Final Check List

目录号 Catalog Number	GSP020002	规格 Specifications	2ml
批号 Lot Number	220929-068-A	有效期 Expiration Date	2025-09-29
生产日期 Manufacture Date	2022-09-29	检验日期 Inspection Date	2022-10-05

检验项目及结果 Inspection Item and Result

检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收:退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品规格、色环颜色正确,管体透明无杂色、晶点、破裂、明显重影;无堵嘴、无严重歪头、歪嘴、刮花磨损等;包装完整, With right specifications and color ring, transparent body. Without crystal point, crack, obvious double shadow. No plugging, deformation, scratches etc. With complete packaging.	依据标准:GB/T 2828.1-2012 Standard:GB/T2828.1-2012 抽样水平:Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
性能 Function	容量允差:±2% Capacity: within ±2%	C=0	10	0:1	符合/Meet criteria	合格/OK
	管体的密封性 Sealing property	C=0	10	0:1	符合/Meet criteria	合格/OK
	油墨牢固度 Ink Fastness	C=0	10	0:1	符合/Meet criteria	合格/OK
	管头超声强度 Ultrasonic intensity of top end	C=0	10	0:1	符合/Meet criteria	合格/OK
	流出时间 Outflow	C=0	10	0:1	符合/Meet criteria	合格/OK
生物性能 Biological Performance	无菌 Sterilized(SAL=10 ⁻⁶)	C=0	3	0:1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	热原≤0.5EU/mL Non-Pyrogenic (≤0.5EU/mL)	C=0	3	0:1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	细胞毒性 Cytotoxicity	C=0	3	0:1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	无酶 DNase & RNase Free	C=0	3	0:1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
最终判断 Final Judgement		接收 Accept ☒ 拒收 Reject ☐				

签名 Signature

检验 Inspected by

审核 Approved by

张 江 强