



巴斯德吸管最终质量检查表 Transfer Pipets Quality Final Check List

目录号 Catalog Number	PP000030	规格 Specifications	3ml			
批号 Lot Number	230802-133-A	有效期 Expiration Date	2028-08-02			
生产日期 Manufacture Date	2023-08-02	检验日期 Inspection Date	2023-08-05			
检验项目及结果 Inspection Item and Result						
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象, 无披锋, 包装完整 No heterochromatic impurities. No oil, crack, shrinkage deformation, burrs. With complete packaging	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
性能 Function	容量: $\pm 5\%$ Capacity: within $\pm 5\%$	C=0	10	0: 1	符合/Meet criteria	合格/OK
生物性能 Biological Performance	无菌 Sterilized(SAL= 10^{-6})	N/A				
最终判断 Final Judgement		接收 Accept	√		拒收 Reject	
签名 Signature						
检验 Inspected by		审核 Approved by				
张 卫 强		马 萧 萧				

