

Quality Certificate

Product Name: 0.2ml PCR Tubes

Catalog Number: PCR410200

Batch Number: 230417-894-F

Customer Focus:

At Jet, our customers' satisfaction is exactly what we are pursuing. We have good management and product development teams who have years of industry experience. Our cell and tissue culture professionals and our entrepreneurial spirit make us responsive to customer needs. We welcome special requests and work to provide the best and most personal service in the industry.

Satisfaction Guarantee:

Jet offers a total satisfaction guarantee so you can be confident in your purchasing decision. If for any reason you are not satisfied with the product performance or service provided by Jet we will either replace or issue a refund for the purchase price of your product.

JET products are manufactured in accordance with ISO 13485: 2016 & ISO 9001: 2015 quality management systems in a class 100,000 clean room.

Non-Pyrogenic - Products labeled non-pyrogenic have been

validated per FDA guidelines on LAL (Limulus Amebocyte

Lysate) testing for medical devices and company guidelines.

The acceptance level for product is less than 0.5 EU/ml.

DNase & RNase Free - This product has been tested and is free of any detectable DNase / RNase contamination.

Human DNA - This product has been tested and free of any detectable Human DNA.

PCR Inhibitor - Products are tested via PCR amplification and gel electrophoresis analysis. Samples must show normal amplification to be considered free of PCR inhibitors.

Quality Control Testing - This product has been inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

Xiao Xiao Ma
Quality Director