

一次性 PCR 单管最终质量检查表 Certificate of Analysis

目录号/批号 Catalog Number/Lot Number	PCR410200/ 230921-1894-F	检验日期 Inspection Date	2023-09-24
生产日期 Manufacture Date	2023-09-21	有效期 Expiration Date	2028-09-21
产品描述 Product Description	0.2ml,PCR 单管,带翻盖,单支,平盖,无 DNA/RNA 酶,不灭菌		
检验项目及标准 Inspection Item and Standard			检验结果 Inspection Result
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象,无披锋,包装完整 No heterochromatic impurities, No oil, crack, shrinkage deformation, burrs. With complete packaging.		☒ 合格/PASS ☐ 不合格/Failure
性能 Function	蒸发率: Evaporation rates PCR 管 (PCR Tubes) ≤ 3%		☒ 合格/PASS ☐ 不合格/Failure
	无菌 Sterility (SAL=10 ⁻⁶)		☒ 合格/PASS ☐ 不合格/Failure
	内毒素 ≤ 0.03EU/mL (无热原) Endotoxin ≤ 0.03EU/mL (Non-Pyrogenic)		☒ 合格/PASS ☐ 不合格/Failure
	无酶 DNase & RNase Free		☒ 合格/PASS ☐ 不合格/Failure
	无抑制剂 PCR Inhibitors Free		☒ 合格/PASS ☐ 不合格/Failure
	无人源 DNA Human DNA Free		☒ 合格/PASS ☐ 不合格/Failure
最终判断 Final Judgement	接收 Accept ☒	拒收 Reject ☐	
签名 Signature			
检验 Inspected by		审核 Approved by	
张仁强		马萧萧	