

### PCR 管最终质量检查表 PCR Tubes Quality Final Check list

<b>目录号</b> Catalog Number	PCR420200	<b>规格</b> Specifications	0.2ml
<b>批号</b> Lot Number	230420-895-A	<b>有效期</b> Expiration Date	2028-04-20
<b>生产日期</b> Manufacture Date	2023-04-20	<b>检验日期</b> Inspection Date	2023-04-23

**检验项目及结果**

**Inspection Item and Result**

检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
<b>外观</b> Appearance	产品无异色杂质、无油污、破裂、缩水变形现象, 无披锋, 包装完整 No heterochromatic impurities, No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
<b>性能</b> Function	蒸发率: Evaporation rates PCR 管 (PCR Tubes) ≤3%	C=0	10	0: 1	符合/Meet criteria	合格/OK
<b>生物性能</b> Biological Performance	无菌 Sterilized(SAL=10 <sup>-6</sup> )	N/A				
	热原≤0.5EU/mL Non-Pyrogenic (≤0.5EU/mL)	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK
<b>最终判断</b> Final Judgement		<b>接收</b> Accept			<b>拒收</b> Reject	

**签名**

**Signature**

检验 Inspected by

审核 Approved by

张 江 强

马 萧 萧

