

一次性 PCR 8 连排管最终质量检查表 Certificate of Analysis

目录号/批号 Catalog Number/Lot Number	PCR421200/230630-895-A	检验日期 Inspection Date	2023-07-03
生产日期 Manufacture Date	2023-06-30	有效期 Expiration Date	2026-06-30
产品描述 Product Description	0.2ml,PCR 8 连排管,排盖,无 DNA/RNA 酶,灭菌		
检验项目及标准 Inspection Item and Standard		检验结果 Inspection Result	
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象,无披锋,包装完整 No heterochromatic impurities, No oil, crack, shrinkage deformation, burrs. With complete packaging.	<input checked="" type="checkbox"/> 合格/PASS <input type="checkbox"/> 不合格/Failure	
性能 Function	蒸发率: Evaporation rates PCR 管 (PCR Tubes) ≤ 3%	<input checked="" type="checkbox"/> 合格/PASS <input type="checkbox"/> 不合格/Failure	
	无菌 Sterility (SAL=10 ⁻⁶)	<input checked="" type="checkbox"/> 合格/PASS <input type="checkbox"/> 不合格/Failure	
	内毒素 ≤ 0.03EU/mL (无热原) Endotoxin ≤ 0.03EU/mL (Non-Pyrogenic)	<input checked="" type="checkbox"/> 合格/PASS <input type="checkbox"/> 不合格/Failure	
	无酶 DNase & RNase Free	<input checked="" type="checkbox"/> 合格/PASS <input type="checkbox"/> 不合格/Failure	
	无抑制剂 PCR Inhibitors Free	<input checked="" type="checkbox"/> 合格/PASS <input type="checkbox"/> 不合格/Failure	
	无人源 DNA Human DNA Free	<input checked="" type="checkbox"/> 合格/PASS <input type="checkbox"/> 不合格/Failure	
最终判断 Final Judgement	接收 Accept <input checked="" type="checkbox"/>	拒收 Reject <input type="checkbox"/>	
签名 Signature			
检验 Inspected by		审核 Approved by	
张记强		马萧萧	

