

一次性 PCR 板最终质量检查表 Certificate of Analysis

目录号/批号 Catalog Number/Lot Number	PCR420096/230428-898-F	检验日期 Inspection Date	2023-04-30
生产日期 Manufacture Date	2023-04-28	有效期 Expiration Date	2028-04-28
产品描述 Product Description	0.2ml 96 孔 PCR 全裙边,透明,不灭菌		
检验项目及标准 Inspection Item and Standard			检验结果 Inspection Result
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象,无披锋,包装完整 No heterochromatic impurities, No oil, crack, shrinkage deformation, burrs. With complete packaging.		☒ 合格/PASS ☐ 不合格/Failure
性能 Function	蒸发率: Evaporation rates PCR 板 (PCR Plates) $\leq 10\%$		☒ 合格/PASS ☐ 不合格/Failure
	无菌 (灭菌产品) Sterility (Only for Sterilized products, SAL= 10^{-6})		☒ 合格/PASS ☐ 不合格/Failure
	内毒素 $\leq 0.03\text{EU/mL}$ (无热原) Endotoxin $\leq 0.03\text{EU/mL}$ (Non-Pyrogenic)		☒ 合格/PASS ☐ 不合格/Failure
	无酶 DNase & RNase Free		☒ 合格/PASS ☐ 不合格/Failure
	无抑制剂 PCR Inhibitors Free		☒ 合格/PASS ☐ 不合格/Failure
	无人源 DNA Human DNA Free		☒ 合格/PASS ☐ 不合格/Failure
最终判断 Final Judgement	接收 Accept ☒	拒收 Reject ☐	
签名 Signature			
检验 Inspected by		审核 Approved by	
张仁强		马萧萧	