

PCR 管最终质量检查表 PCR Tubes Quality Final Check list

目录号 Catalog Number		PCR500096		规格 Specifications		0.2ml	
批号 Lot Number		230619-916-F		有效期 Expiration Date		2026-06-19	
生产日期 Manufacture Date		2023-06-19		检验日期 Inspection Date		2023-06-22	
检验项目及结果 Inspection Item and Result							
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result		
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac: Re	结果 Result	判断 Judgement	
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象, 无披锋, 包装完整 No heterochromatic impurities, No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria		合格/OK
性能 Function	蒸发率: Evaporation rates PCR 管 (PCR Tubes) ≤3%	C=0	10	0: 1	符合/Meet criteria		合格/OK
生物性能 Biological Performance	无菌 Sterilized(SAL=10 ⁻⁶)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -		合格/OK
	热原≤0.03EU/mL Non-Pyrogenic (≤0.03EU/mL)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -		合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -		合格/OK
最终判断 Final Judgement		接收 Accept			拒收 Reject		
签名 Signature							
检验 Inspected by				审核 Approved by			
张 江 伟				马 萧 萧			