



培养液瓶最终质量检查报告 Solution Bottles Quality Final Check List						
目录号 Catalog Number		CTF010150		规格 Specifications		150mL
批号 Lot Number		230616-109-A		有效期 Expiration Date		2026-06-16
生产日期 Manufacture Date		2023-06-16		检验日期 Inspection Date		2023-06-19
检验项目及结果 Inspection Item and Result						
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无油污、破裂、缩水变形现象, 无披锋, 包装完整 No oil, crack, shrinkage deformation, burr. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
性能 Function	容量: 允差: $\pm 5\%$ Capacity: within $\pm 5\%$	C=0	10	0:1	符合/Meet criteria	合格/OK
	耐温性能: -20°C, 24h; Temperature resistance : -20°C , 24 hour	C=0	10	0:1	符合/Meet criteria	合格/OK
生物性能 Biological Performance	无菌 Sterilized(SAL= 10^{-6})	C=0	3	0:1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	热原 ≤ 0.03 EU/mL Non-Pyrogenic (≤ 0.03 EU/mL)	C=0	3	0:1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	细胞毒性 Cytotoxicity	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
最终判断 Final Judgement		接收 Accept	√		拒收 Reject	
签名 Signature						
检验 Inspected by		审核 Approved by				
张 江 强		马 菊 菊				