



广州华大生物科技有限公司
Guangzhou Huada Biotechnology Co., Ltd.

辐照证明书

CERTIFICATE OF IRRADIATION

2127235

HD/QP/20-03 V3.1

单位名称: 广州洁特生物过滤股份有限公司

CompanyName:

产品名称	目录号	批号	数量
Article Description	Catalog No.	Lot No.	Carton Qty(cs)
70µm Cell Strainer	CSS013070	221223-300-A	38
250ml Solution Bottles	CTF010250	221213-110-A	17
2ml Serological Pipets	EPS02N	221220-296-A	1
6cm Tissue Culture Dishes	ET2060	221220-206BA	11
6cm Tissue Culture Dishes	ET2060	221220-206BA	2
48well Tissue Culture Plates	ET3048	221224-077-F	10
48well Tissue Culture Plates	ET3048	221224-077-F	1

合同号: W000200653

生产单号: P000195463

辐照结束日期: 2022/12/25

CONTRACT No:

IRRADIATION LOT No:

IRRADIATION LOT FINISH
DATE:

这批产品已经过电子束射线辐照。

THE PRODUCTS HAVE BEEN IRRADIATED BY ELECTRON BEAM.

客户要求吸收剂量: 15.0kGy-30.0kGy

产品实际吸收剂量: 16.8kGy-23.0kGy

SPECIFIED DOSE:

DELIVERED DOSE:

本公司兹证明以上所列产品, 在吸收剂量测试系统精度和准确度的使用范围内接受到如上产品实际吸收剂量。Our company certifies that the products listed above(as described by the manufacture) received the delivered doses within the precision and accuracy of the dosimetry system employed.

检验员: 戴绮珊

审核: 王燕芳

日期: 2022/12/25

INSPECTOR:

APPROVED:

DATE:

本证明书复印件无效

THE COPY OF THIS CERTIFICATE IS INVALID

本公司已通过TÜV SÜD 公司的 ISO9001: 2015 & ISO13485: 2016管理体系认证


Our company has been granted certificates of ISO9001:2015&ISO13485:2016 by TÜV SÜD company

地址:(Add):广州市萝岗区科学城开泰大道伴河路84号

电话(TEL):+86-20-32203400 32203401 82099609

传真(Fax):+86-20-82099662

网址 (website):www.huadashengwu.com





广州华大生物科技有限公司
Guangzhou Huada Biotechnology Co., Ltd.

辐照证明书

CERTIFICATE OF IRRADIATION

2126653

HD/QP/20-03 V3.1

单位名称：广州洁特生物过滤股份有限公司

CompanyName:

产品名称	目录号	批号	数量
Article Description	Catalog No.	Lot No.	Carton Qty(cs)
50ml Centrifuge Tubes	RED 5025B	221119-060-AA	69
1000ml Square Media Bottles	SSB010000	221117-858-F	5
1000ml Square Media Bottles	SSB010000-ZX	221117-858-F	5
250ml Square Media Bottles	SSB010250	221109-865-F	33
500ml Square Media Bottles	SSB130500-ZX	221117-807-F	5
250ml Erlenmeyer Flasks	TAB012250	221112-756-A	10
1000ml Erlenmeyer Flasks	TAB102000	221118-719-A	2

合同号：W000198954

生产单号：P000193319

辐照结束日期：2022/11/21

CONTRACT No:

IRRADIATION LOT No:

IRRADIATION LOT FINISH
DATE:

这批产品已经过电子束射线辐照。

THE PRODUCTS HAVE BEEN IRRADIATED BY ELECTRON BEAM.

客户要求吸收剂量：15.0kGy-30.0kGy

产品实际吸收剂量：16.8kGy-23.0kGy

SPECIFIED DOSE:

DELIVERED DOSE:

本公司兹证明以上所列产品，在吸收剂量测试系统精度和准确度的使用范围内接受到如上产品实际吸收剂量。Our company certifies that the products listed above(as described by the manufacture) received the delivered doses within the precision and accuracy of the dosimetry system employed.

检验员：戴绮珊

审核：王燕芳

日期：2022/11/21

INSPECTOR:

APPROVED:

DATE:

盖章：

SEAL: 质量检验专用章

本证明书复印件无效

THE COPY OF THIS CERTIFICATE IS INVALID

本公司已通过TÜV SÜD 公司的 ISO9001: 2015 & ISO13485: 2016管理体系认证

Our company has been granted certificates of ISO9001:2015&ISO13485:2016 by TÜV SÜD company

地址:(Add):广州市萝岗区科学城开泰大道伴河路84号

电话(TEL):+86-20-32203400 32203401 82099609

传真(Fax):+86-20-82099662

网址 (website):www.huadashengwu.com



广州华大生物科技有限公司
Guangzhou Huada Biotechnology Co., Ltd.

辐照证明书

CERTIFICATE OF IRRADIATION

2126492

HD/QP/20-03 V3.1

单位名称：广州洁特生物过滤股份有限公司
CompanyName:

产品名称	目录号	批号	数量
Article Description	Catalog No.	Lot No.	Carton Qty(cs)
10ml Aspirating Pipets	GSP100010	221114-070-A	45
100ml Serological Pipets	GSP110100	221113-245-A	41
100ml Serological Pipets	GSP110100	221113-245-A	20
10ml Serological Pipets	GSP120010	221111-070-A	45
6well Tissue Culture Plates	I.122.11.006	221113-074-F	10
6well Tissue Culture Plates	I.122.11.006	221113-074-F	1
96well Tissue Culture Plates	I.122.11.096	221114-078-F	27

合同号：W000198636

生产单号：P000192850

辐照结束日期：2022/11/15

CONTRACT No:

IRRADIATION LOT No:

IRRADIATION LOT FINISH
DATE:

这批产品已经过电子束射线辐照。

THE PRODUCTS HAVE BEEN IRRADIATED BY ELECTRON BEAM.

客户要求吸收剂量：15.0kGy-30.0kGy
SPECIFIED DOSE:

产品实际吸收剂量：16.7kGy-22.7kGy
DELIVERED DOSE:

本公司兹证明以上所列产品，在吸收剂量测试系统精度和准确度的使用范围内接受到如上产品实际吸收剂量。Our company certifies that the products listed above(as described by the manufacture) received the delivered doses within the precision and accuracy of the dosimetry system employed.

检验员：戴绮珊

审核：王燕芳

日期：2022/11/15

INSPECTOR:

APPROVED:

DATE:

本证明书复印件无效

THE COPY OF THIS CERTIFICATE IS INVALID

本公司已通过TÜV SÜD 公司的 ISO9001: 2015 & ISO13485: 2016管理体系认证

Our company has been granted certificates of ISO9001:2015&ISO13485:2016 by TÜV SÜD company

地址(Add):广州市萝岗区科学城开泰大道伴河路84号

电话(TEL):+86-20-32203400 32203401 82099609

传真(Fax):+86-20-82099662

网址 (website):www.huadashengwu.com





广州华大生物科技有限公司
Guangzhou Huada Biotechnology Co., Ltd.

辐照证明书

2125971

CERTIFICATE OF IRRADIATION

HD/QP/20-03 V3.1

单位名称: 广州洁特生物过滤股份有限公司

Company Name:

产品名称	目录号	批号	数量
Article Description	Catalog No.	Lot No.	Carton Qty(cs)
50ml Centrifuge Tubes	CFT011500	221012-060-AA	136
50ml Centrifuge Tubes	CFT011500	221013-060-AA	90
225ml Conical Centrifuge Tubes	CFT012225	221014-215-A	1
5ml Lid Lock Micro centrifuge Tubes	CFT322050	221013-340-D	5
50ml Centrifuge Tubes	CFT511500	221014-060-D	73
50ml Centrifuge Tubes	CFT521500	221014-060-D	8
50ml Centrifuge Tubes	CFT621500	221014-060-D	5

合同号: W000197169

生产单号: P000191228

辐照结束日期: 2022/10/18

CONTRACT No:

IRRADIATION LOT No:

IRRADIATION LOT FINISH
DATE:

这批产品已经过电子束射线辐照。

THE PRODUCTS HAVE BEEN IRRADIATED BY ELECTRON BEAM.

客户要求吸收剂量: 15.0kGy-30.0kGy

产品实际吸收剂量: 16.8kGy-23.0kGy

SPECIFIED DOSE:

DELIVERED DOSE:

本公司兹证明以上所列产品, 在吸收剂量测试系统精度和准确度的使用范围内接受到如上产品实际吸收剂量。Our company certifies that the products listed above(as described by the manufacture) received the delivered doses within the precision and accuracy of the dosimetry system employed.

检验员: 戴绮珊

审核: 王燕芳

日期: 2022/10/18

INSPECTOR:

APPROVED:

DATE:

本证明书复印件无效

THE COPY OF THIS CERTIFICATE IS INVALID

本公司已通过TÜV SÜD 公司的 ISO9001: 2015 & ISO13485: 2016管理体系认证

Our company has been granted certificates of ISO9001:2015&ISO13485:2016 by TÜV SÜD company

地址(Add): 广州市萝岗区科学城开泰大道伴河路84号

电话(TEL): +86-20-32203400 32203401 82099609

传真(Fax): +86-20-82099662

网址 (website): www.huadashengwu.com





广州华大生物科技有限公司
Guangzhou Huada Biotechnology Co., Ltd.

辐照证明书

CERTIFICATE OF IRRADIATION

2126419

HD/QP/20-03 V3.1

单位名称: 广州洁特生物过滤股份有限公司
Company Name:

产品名称	目录号	批号	数量
Article Description	Catalog No.	Lot No.	Carton Qty (cs)
30mm Syringe Driven Filters	SD Filter-PES-0/45-30-S	221106-052-A	1
13mm Syringe Driven Filters	SD Filter-PVDF-0/45-13-S	221106-051-A	1
30mm Syringe Driven Filters	SD Filter-PVDF-0/45-30-S	221109-052-A	4
250ml Square Media Bottles	SSB010250	221109-865-F	29
250ml Square Media Bottles	SSB010250	221109-865-F	4
5ml Serum&Sample tubes	SST001050	221103-131-D	1
125ml Erlenmeyer Flasks	TAB002125	221108-686-F	42

合同号: W000198512

生产单号: P000192712

辐照结束日期: 2022/11/12

CONTRACT No:

IRRADIATION LOT No:

IRRADIATION LOT FINISH
DATE:

这批产品已经过电子束射线辐照。

THE PRODUCTS HAVE BEEN IRRADIATED BY ELECTRON BEAM.

客户要求吸收剂量: 15.0kGy-30.0kGy
SPECIFIED DOSE:

产品实际吸收剂量: 16.8kGy-23.0kGy
DELIVERED DOSE:

本公司兹证明以上所列产品, 在吸收剂量测试系统精度和准确度的使用范围内接受到如上产品实际吸收剂量。Our company certifies that the products listed above (as described by the manufacturer) received the delivered doses within the precision and accuracy of the dosimetry system employed.

检验员: 戴绮珊

审核: 王燕芳

日期: 2022/11/12

INSPECTOR:

APPROVED:

DATE:

本证明书复印件无效

THE COPY OF THIS CERTIFICATE IS INVALID

本公司已通过TÜV SÜD 公司的 ISO9001: 2015 & ISO13485: 2016管理体系认证


Our company has been granted certificates of ISO9001:2015&ISO13485:2016 by TÜV SÜD company

地址:(Add): 广州市萝岗区科学城开泰大道伴河路84号

电话(TEL): +86-20-32203400 32203401 82099609

传真(Fax): +86-20-82099662

网址 (website): www.huadashengwu.com





广州华大生物科技有限公司
Guangzhou Huada Biotechnology Co., Ltd.

辐照证明书

CERTIFICATE OF IRRADIATION

2126324

HD/QP/20-03 V3.1

单位名称：广州洁特生物过滤股份有限公司

Company Name:

产品名称	目录号	批号	数量
Article Description	Catalog No.	Lot No.	Carton Qty (cs)
100~1000 μ l Pipette Micro Tips	PPT371000	221106-598-D	24
10 μ l Pipette Micro Tips	PPT451010	221105-361-D	224
6cm Tissue Culture Dishes	sc-358783	221030-206AA	7
13mm Syringe Driven Filters	SD Filter-NYLON-0/22-13-S	221103-051-A	5
13mm Syringe Driven Filters	SD Filter-PES-0/22-13-S	221030-051-A	20
13mm Syringe Driven Filters	SD Filter-PES-0/22-13-S	221030-051-A	2
125ml Square Media Bottles	SSB010125	221103-808-F	50

合同号：W000198242

生产单号：P000192320

辐照结束日期：2022/11/7

CONTRACT No:

IRRADIATION LOT No:

IRRADIATION LOT FINISH
DATE:

这批产品已经过电子束射线辐照。

THE PRODUCTS HAVE BEEN IRRADIATED BY ELECTRON BEAM.

客户要求吸收剂量：15.0kGy-30.0kGy

产品实际吸收剂量：16.7kGy-22.9kGy

SPECIFIED DOSE:

DELIVERED DOSE:

本公司兹证明以上所列产品，在吸收剂量测试系统精度和准确度的使用范围内接受到如上产品实际吸收剂量。Our company certifies that the products listed above (as described by the manufacture) received the delivered doses within the precision and accuracy of the dosimetry system employed.

检验员：戴绮珊

审核：王燕芳

日期：2022/11/7

INSPECTOR:

APPROVED:

DATE:

本证明书复印件无效

THE COPY OF THIS CERTIFICATE IS INVALID

本公司已通过TÜV SÜD 公司的 ISO9001: 2015 & ISO13485: 2016管理体系认证


Our company has been granted certificates of ISO9001:2015&ISO13485:2016 by TÜV SÜD company

地址:(Add):广州市萝岗区科学城开泰大道伴河路84号

电话(TEL):+86-20-32203400 32203401 82099609

传真(Fax):+86-20-82099662

网址 (website):www.huadashengwu.com





广州华大生物科技有限公司
Guangzhou Huada Biotechnology Co., Ltd.

辐照证明书

CERTIFICATE OF IRRADIATION

2126289

HD/QP/20-03 V3.1

单位名称: 广州洁特生物过滤股份有限公司

CompanyName:

产品名称	目录号	批号	数量
Article Description	Catalog No.	Lot No.	Carton Qty(cs)
10ml Serological Pipets	GSP110010	221103-070-A	220
10ml Serological Pipets	GSP110010	221103-070-A	17
10ml Serological Pipets	GSP110010	221103-070-A	5
10ml Serological Pipets	GSP110010	221103-070-A	88
25ml Serological Pipets	GSP120025	221103-071-A	36
L-Shape Cell Spreader	LG-CSP012014S	221030-261-D	5
24well Tissue Culture Plates	LG-TCP011024S	221103-076-F	20

合同号: W000198159

生产单号: P000192318

辐照结束日期: 2022/11/5

CONTRACT No:

IRRADIATION LOT No:

IRRADIATION LOT FINISH
DATE:

这批产品已经过电子束射线辐照。

THE PRODUCTS HAVE BEEN IRRADIATED BY ELECTRON BEAM.

客户要求吸收剂量: 15.0kGy-30.0kGy

产品实际吸收剂量: 16.8kGy-23.0kGy

SPECIFIED DOSE:

DELIVERED DOSE:

本公司兹证明以上所列产品, 在吸收剂量测试系统精度和准确度的使用范围内接受到如上产品实际吸收剂量。Our company certifies that the products listed above(as described by the manufacture) received the delivered doses within the precision and accuracy of the dosimetry system employed.

检验员: 戴绮珊

审核: 王燕芳

日期: 2022/11/5

INSPECTOR:

APPROVED:

DATE:

本证明书复印件无效

THE COPY OF THIS CERTIFICATE IS INVALID



本公司已通过TÜV SÜD 公司的 ISO9001: 2015 & ISO13485: 2016管理体系认证

Our company has been granted certificates of ISO9001:2015&ISO13485:2016 by TÜV SÜD company

地址:(Add):广州市萝岗区科学城开泰大道伴河路84号

电话(TEL):+86-20-32203400 32203401 82099609

传真(Fax):+86-20-82099662

网址 (website):www.huadashengwu.com