

加样槽最终质量检查表 Liquid Transfer Trough Quality Final Check list

目录号 Catalog Number	LTT052050	规格 Specifications	50ml			
批号 Lot Number	220409-336-D	有效期 Expiration Date	2025-04-09			
生产日期 Manufacture Date	2022-04-09	检验日期 Inspection Date	2022-04-12			
检验项目及结果 Inspection Item and Result						
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收:退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象,无披锋,包装完整 No heterochromatic impurities. No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
生物性能 Biological Performance	无菌 Sterilized (SAL=10 ⁻⁶)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -	合格/OK
最终判断 Final Judgement		接收 Accept	√		拒收 Reject	
签名 Signature						
检验 Inspected by		审核 Approved by				
张 江 强		马 菊 菊				

