

吸头最终质量检查表 Tips Quality Final Check list

目录号 Catalog Number	PMT010000	规格 Specifications	1000μl (非灭菌)
批号 Lot Number	220528-307-D	有效期 Expiration Date	2027-05-28
生产日期 Manufacture Date	2022-05-28	检验日期 Inspection Date	2022-05-31

检验项目及结果

Inspection Item and Result

检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无油污、破裂、缩水变形现象，无披锋；内外壁光滑，无弯曲；包装完整 No oil, crack, shrinkage deformation, burrs. The tips is smooth, without bending. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
性能 Function	与移液枪匹配性: Match with pipettor	C=0	10	0: 1	符合/Meet criteria	合格/OK
	管壁残留 Liquid residue on the inner wall of the tip	C=0	10	0: 1	符合/Meet criteria	合格/OK
	耐高温: 121℃, 20 分钟 High temperature performance: 121℃, 20min	N/A				
生物性能 Biological Performance	无菌 Sterilized (SAL=10 ⁻⁶)	N/A				
	热原≤0.5EU/mL Non-Pyrogenic (≤0.5EU/mL)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -	合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -	合格/OK
最终判断 Final Judgement		接收 Accept	√		拒收 Reject	
签名 Signature						
检验 Inspected by		审核 Approved by				
张江伟		马荣荣				