

## 酶标板最终质量检查表 ELISA Products Quality Final Check list

|                          |              |                         |            |
|--------------------------|--------------|-------------------------|------------|
| 目录号<br>Catalog Number    | FEP101896    | 规格<br>Specifications    | 96 孔       |
| 批号<br>Lot Number         | 221208-081-A | 有效期<br>Expiration Date  | 2025-12-08 |
| 生产日期<br>Manufacture Date | 2022-12-08   | 检验日期<br>Inspection Date | 2022-12-11 |

### 检验项目及结果

#### Inspection Item and Result

| 检验项目及标准<br>Inspection Item and Standard |                                                                                                                                        | 接收水平<br>Acceptance Level                                                                                          |                    |               | 检验结果<br>Inspection Result |                 |
|-----------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------|--------------------|---------------|---------------------------|-----------------|
|                                         |                                                                                                                                        | 检验水平<br>Quality Level                                                                                             | 样品数<br>Sample Size | 收: 退<br>Ac:Re | 结果<br>Result              | 判断<br>Judgement |
| 外观<br>Appearance                        | 产品无异色杂质、无油污、破裂、缩水变形现象，无披锋；包装完整<br>No heterochromatic impurities.No oil, crack, shrinkage, deformation, burrs. With complete packaging. | 依据标准: GB/T 2828.1-2012<br>Standard: GB/T2828.1-2012<br>抽样水平: Level II, AQL 1.5<br>Sample Level: Level II, AQL 1.5 |                    |               | 符合/Meet criteria          | 合格/OK           |
| 性能<br>Function                          | CV 测试: 整板 CV 值<5%，板间 CV 值 < 10%<br>CV Test:<br>CV of single plate<5%<br>CV of different plates < 10%                                   | C=0                                                                                                               | 1                  | 0: 1          | 符合/Meet criteria          | 合格/OK           |
|                                         | OD 值测试: 各孔 OD > 0.8<br>OD Test:<br>OD of Each hole > 0.8                                                                               | C=0                                                                                                               | 1                  | 0: 1          | 符合/Meet criteria          | 合格/OK           |
|                                         | 蛋白吸附: 300~400ng/cm3<br>protein adsorption                                                                                              | C=0                                                                                                               | 1                  | 0: 1          | 符合/Meet criteria          | 合格/OK           |
| 最终判断<br>Final Judgement                 |                                                                                                                                        | 接收<br>Accept                                                                                                      |                    |               | 拒收<br>Reject              |                 |
| 签名<br>Signature                         |                                                                                                                                        |                                                                                                                   |                    |               |                           |                 |
| 检验 Inspected by                         |                                                                                                                                        | 审核 Approved by                                                                                                    |                    |               |                           |                 |
| 张江强                                     |                                                                                                                                        | 马荣荣                                                                                                               |                    |               |                           |                 |