

酶标板最终质量检查表 ELISA Products Quality Final Check list

目录号 Catalog Number		FEP101896		规格 Specifications		96 孔	
批号 Lot Number		230523-081-A		有效期 Expiration Date		2026-05-23	
生产日期 Manufacture Date		2023-05-23		检验日期 Inspection Date		2023-05-26	
检验项目及结果 Inspection Item and Result							
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result		
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement	
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象, 无披锋; 包装完整 No heterochromatic impurities.No oil, crack, shrinkage, deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK	
	CV 测试: 整板 CV 值 < 5%, 板间 CV 值 < 10% CV Test: CV of single plate < 5% CV of different plates < 10%	C=0	1	0: 1	符合/Meet criteria	合格/OK	
	OD 值测试: 各孔 OD > 0.8 OD Test: OD of Each hole > 0.8	C=0	1	0: 1	符合/Meet criteria	合格/OK	
性能 Function	蛋白吸附: 300~400ng/cm3 protein adsorption	C=0	1	0: 1	符合/Meet criteria	合格/OK	
最终判断 Final Judgement		接收 Accept			拒收 Reject		
签名 Signature							
检验 Inspected by				审核 Approved by			
张 江 华				马 萧 萧			