



Quality Certificate

Product Name: 96 well ELISA Plates

Catalog Number : FEP101296 Batch Number : 210901-081

Customer Focus:

At Jet, our customers' satisfaction is exactly what we are pursuing. We have good management and product development teams who have years of industry experience. Our cell and tissue culture professionals and our entrepreneurial spirit make us responsive to customer needs. We welcome special requests and work to provide the best and most personal service in the industry.

Satisfaction Guarantee:

Jet offers a total satisfaction guarantee so you can be confident in your purchasing decision. If for any reason you are not satisfied with the product performance or service provided by Jet, we will either replace or issue a refund for the purchase price of your product.

JET products are manufactured in accordance with ISO 13485: 2016 & ISO 9001: 2015 quality management systems in a class 100,000 clean room.

Binding Capacity - Products labeled as high binding have been processed and validated. Each lot has been taken some samples random from total Lot as representative sample for actual binding capacity test and they comply with our product specification.

CV of Transmittance - Each lot has been taken some samples random from total Lot as representative sample for CV value and they comply with our product specification. The CV of transmittance of wells is not more than 5%, and the CV of transmittance of plates is not more than 10%.

Non-pyrogenic - Products labeled non-pyrogenic have been validated per FDA guidelines on LAL (Limulus Amebocyte Lysate) testing for medical devices and company guidelines. The acceptance level for product is less than 0.5 EU/device.

Quality Control Testing - Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

Quality Manager