

超滤离心管最终质量检查表 Centrifugal Filters Quality Final Check List

目录号 Catalog Number	FTT450500	规格 Specifications	50ml; 50KD
批号 Lot Number	230512-880-F	有效期 Expiration Date	2028-05-12
生产日期 Manufacture Date	2023-05-12	检验日期 Inspection Date	2023-05-15

检验项目及结果

Inspection Item and Result

检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象, 无披锋, 包装完整 No heterochromatic impurities. No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
性能 Function	离心力测试: 摆桶式:3000×g 定角式:4000×g Centrifuge test : Swing bucket rotor :3000×g Fixed angle rotor: 4000×g	C=0	4	0: 1	符合/Meet criteria	合格/OK
	蛋白截留率: > 80% Protein retention rate: > 80%	C=0	4	0: 1	符合/Meet criteria	合格/OK
	容量: ±2% Capacity: within±2%	C=0	4	0: 1	符合/Meet criteria	合格/OK
	油墨牢固度 Ink fastness	C=0	4	0: 1	符合/Meet criteria	合格/OK
	配合密封 Fit and sealing	C=0	4	0: 1	符合/Meet criteria	合格/OK
	书写效果 writing effect	C=0	4	0: 1	符合/Meet criteria	合格/OK
最终判断 Final Judgement		接收 Accept	√		拒收 Reject	

签名

Signature

检验 Inspected by

审核 Approved by

张 江 强



马 萧 萧