

U 型板最终质量检查表 U-Well Plates Quality Final Check list

目录号 Catalog Number	UWP043096	规格 Specifications	96 U-Well			
批号 Lot Number	230919-562-A	有效期 Expiration Date	2026-09-19			
生产日期 Manufacture Date	2023-09-19	检验日期 Inspection Date	2023-09-22			
检验项目及结果 Inspection Item and Result						
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level		检验结果 Inspection Result		
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品粘结部牢固, 无气泡、条纹, 变形, 无漏液现象等, 包装完整 Firm bonding part of the product, no bubbles, stripes, deformation, no leakage phenomenon, etc. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
生物性能 Biological Performance	无菌 Sterilized (SAL=10 ⁻⁶)	C=0	10	0: 1	样品 Samples: --- 阳性对照 Positive: +++ 阴性对照 Negative: ---	合格/OK
最终判断 Final Judgement		接收 Accept	√		拒收 Reject	
签名 Signature						
检验 Inspected by		审核 Approved by				
张 江 强		马 菊 菊				