

发光板最终质量检查表 Luminescence Test Plates Quality Final Check List

目录号 Catalog Number	LTP010296	规格 Specifications	96 孔			
批号 Lot Number	221205-005-A	有效期 Expiration Date	2027-12-05			
生产日期 Manufacture Date	2022-12-05	检验日期 Inspection Date	2022-12-08			
检验项目及结果 Inspection Item and Result						
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level		检验结果 Inspection Result		
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象, 无披锋, 包装完整 No heterochromatic impurities. No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
生物性能 Biological Performance	热原 ≤ 0.5 EU/mL Non-Pyrogenic (≤ 0.5 EU/mL)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -	合格/OK
最终判断 Final Judgement		接收 Accept	√		拒收 Reject	
签名 Signature						
检验 Inspected by		审核 Approved by				
张 记 强		质量部			马 萧 萧	