

发光板最终质量检查表 Luminescence Test Plates Quality Final Check List

目录号 Catalog Number		LTP021896		规格 Specifications		96 孔	
批号 Lot Number		220315-005-A		有效期 Expiration Date		2027-03-15	
生产日期 Manufacture Date		2022-03-15		检验日期 Inspection Date		2022-03-18	
检验项目及结果 Inspection Item and Result							
检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result		
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement	
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象，无披锋，包装完整 No heterochromatic impurities. No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria		合格/OK
生物性能 Biological Performance	热原≤0.5EU/mL Non-Pyrogenic (≤0.5EU/mL)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -		合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: + + + 阴性对照 Negative: - - -		合格/OK
最终判断 Final Judgement		接收 Accept		√		拒收 Reject	
签名 Signature							
检验 Inspected by				审核 Approved by			
张 卫 强				马 菊 菊			