



接种针接种环最终质量检查表 Inoculating Loops & needles Quality Final Check list

目录号 Catalog Number	DIL112010	规格 Specifications	10 µl
批号 Lot Number	230110-268-A	有效期 Expiration Date	2026-01-10
生产日期 Manufacture Date	2023-01-10	检验日期 Inspection Date	2023-01-13

检验项目及结果

Inspection Item and Result

检验项目及标准 Inspection Item and Standard		接收水平 Acceptance Level			检验结果 Inspection Result	
		检验水平 Quality Level	样品数 Sample Size	收: 退 Ac:Re	结果 Result	判断 Judgement
外观 Appearance	产品无异色杂质、无油污、破裂、缩水变形现象, 无披锋, 包装完整 No heterochromatic impurities. No oil, crack, shrinkage deformation, burrs. With complete packaging.	依据标准: GB/T 2828.1-2012 Standard: GB/T2828.1-2012 抽样水平: Level II, AQL 1.5 Sample Level: Level II, AQL 1.5			符合/Meet criteria	合格/OK
生物性能 Biological Performance	无菌 Sterilized (SAL=10 ⁻⁶)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -	合格/OK
	热原≤0.5EU/mL Non-Pyrogenic (≤0.5EU/mL)	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -	合格/OK
	无酶 DNase & RNase Free	C=0	3	0: 1	样品 Samples: - - - 阳性对照 Positive: +++ 阴性对照 Negative: - - -	合格/OK
最终判断 Final Judgement		接收 Accept	√		拒收 Reject	

签名

Signature

检验 Inspected by	审核 Approved by
张 卫 强	马 萧 萧

