Agilent Technologies

ULTRA SCIENTIFIC Analytical Solutions

Certificate of Analysis

FID MDL Sample

Agilent Part Number:	**5188-5372**	**Sample Lot Number:**	**CE-2751**

Concentrations:

n-Tridecane	2.37 mg/L (±2%)	2.0 ng C/µL
n-Tetradecane	2.37 mg/L (±2%)	2.0 ng C/µL
n-Pentadecane	23.6 mg/L (±2%)	20. ng C/µL
n-Hexadecane	23.6 mg/L (±2%)	20. ng C/µL

In isooctane (2,2,4-trimethylpentane) as solvent

Purities:

n-Tridecane	99%
n-Tetradecane	99%
n-Pentadecane	97+%
n-Hexadecane	97+%
Isooctane	99%

This standard has been produced gravimetrically and volumetrically using ISO9001 quality procedures. Balances used are calibrated regularly and meet the requirements of ISO 17025 augmented by ANSI/NCSL Z540-1-1994. Gas chromatography was used to verify the standard concentration.

Typical Analytical Spectrum or Chromatogram

GC Chromatogram – n-tridecane**,** n-tetradecane, n-pentadecane and n-hexadecane in isooctane



Date of release: 08 August, 2008

Date of expiration: 30 September, 2011

John Russo

John Russo
QC Manager
ULTRA Scientific, Inc.