Agilent Technologies

Certificate of Analysis

QC Solution, AOAC Method 2007.01

Agilent Part Number: **5190-0503**

Sample Lot Number: **CS-2721**

This analytical reference material was manufactured and verified in accordance with an ISO 9001 registered quality system, and the analyte concentrations were verified by an ISO 17025 accredited laboratory. The certified value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	Certified Value
triphenyl phosphate	000115-86-6	RM01112	501.8 ± 2.5 ug/mL

Solvent: acetonitrile with trace acetic acid (0.1%)

Calibrated Class A glassware and clean bottles were used in the manufacture of this standard. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Date of release: **17 May 2018**
Date of expiration: **30 June 2020**

John Russo
President

Monica Bourgeois
Director of QA/RA