Agilent Technologies

ULTRA SCIENTIFIC
Analytical Solutions

Certificate of Analysis

QC Solution, AOAC Method 2007.01

Agilent Part Number: **5190-0503**

Sample Lot Number: **CG-0586A**

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA Scientific's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The certified value and uncertainty value for each analyte was determined gravimetrically.

Analyte	CAS#	Analyte Lot	True Value
triphenyl phosphate	000115-86-6	DZ-11408MX	501.5 ± 2.5 ug/mL

Solvent: acetonitrile with trace acetic acid (0.1%)

ULTRA Scientific uses calibrated Class A glassware and clean bottles in the manufacture of ULTRA standards. Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Date of release: **13 May 2011**
Date of expiration: **30 June 2013**

William J. Lea
Quality Assurance Manager