

CERTIFICATE OF ANALYSIS

A52730 **QuantStudio™ Absolute Q™ Isolation Buffer**

Packaging Lot: 3283487

Expiry Date: 20.01.2027 (DD.MM.YYYY)

Storage: at room temperature

Note: **For Research Use Only. Not for use in diagnostic procedures.**

Filling lots for components in package:

Lot	Quantity	Description
3283487	3 mL	Absolute Q™ Isolation Buffer

QUALITY CONTROL

Parameter	Method	Requirement	Result
Appearance (color)	Visual inspection.	Colorless	Pass
Appearance (form)	Visual inspection.	Liquid	Pass
Structure	Infrared Spectrum.	Conforms to Structure	Pass
Volatiles	Volatility measurement.	≤ 1.00 %	Pass
Viscosity	Viscosity measurement at 25° C.	9.0 – 11.0	Pass

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by: **J. Žilinskienė**

