Thermo Fisher Scientific
Nunc A/S
Kamstrupvej 90, P.O.Box 280
DK-4000 Roskilde, Denmark
+45 4631 2000 Tel
Info.nunc@thermofisher.com
VAT/SE NO. DK 39582910
www.thermofisher.com

Certificate of Conformity

Product: Nunc 1 µl Inoculation Loop, PS, Clear

Date: 2022-11-22

Item number: 253287
Lot number: 176744
Date of manufacture: October 2022
Date of expiry: October 2027

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

The loops are manufactured to deliver a volume of 1 µL ± 0.3µL.

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation", with SAL 10E-6.

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality management system - Requirements" and ISO 13485:2016 "Medical devices - Quality management systems - System requirements for regulatory purposes".

Roskilde Site

Anita Kvistgaard
Operations Quality Manager
Please note that this is an electronically generated signature