

MicroSeq D2 LSU rDNA Fungal PCR Kit

Product No. **4349153**
 Lot No. **2412136**
 Date of Manufacture **17DEC2024**
 Expiration Date **16DEC2025**

TEST	SPECIFICATION	RESULT
Positive PCR Amplification Test		
Correct amplification of a 292 bp DNA fragment using 15 ng per 30 uL PCR reaction of total <i>Saccharomyces cerevisiae</i> genomic DNA (positive control) is verified by sequencing of this fragment with the		

The combined forward and reverse sequencing accuracy is at least 99.5 %	> = 99.5 %	100.0 %
---	------------	---------

The combined forward and reverse sequencing clear read is at least 276 base pairs.	> = 276 bp	281 bp
--	------------	--------

Negative PCR Amplification Test		
--	--	--

No match is found on combined forward and reverse sequencing of the PCR product generated when using negative control water (instead of <i>Saccharomyces cerevisiae</i> genomic DNA), indicating that the kit components are free of contamination.		Pass
---	--	------

Physical Inspection		
----------------------------	--	--

Correct label and lot number.		Pass
-------------------------------	--	------

Caps are tight with no signs of leakage.		Pass
--	--	------

For Research Use Only.
Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Warrington, UK



Corporate Legal Entity
 Life Technologies
 5781 Van Allen Way
 Carlsbad, CA, USA 92008
www.thermofisher.com
 For inquiries, contact us at
cofarequests@thermofisher.com

Quality Assurance
Issued 20DEC2024